

NATtrol™ Respiratory Verification Panel 2.1

PRODUCT DESCRIPTION:

NATtrol™ Respiratory Verification Panel 2.1 (qualitative)* is formulated with purified, intact bacterial cells and viral particles. The microorganisms have been chemically modified to render them non-infectious and refrigerator stable. NATRVP2.1-BIO contains 23 x 0.6 mL vials of bacterial and viral NATtrol™ and 1 x 0.6 mL of negative (matrix only) as listed in Table 1. The panel members are supplied in a proprietary matrix.

*Pat: <http://www.zeptometrix.com/patent-information/>

INTENDED USE:

NATtrol™ Respiratory Verification Panel 2.1 can be used to develop procedures for verification of performance of molecular assays that detect the presence of bacterial and viral nucleic acids (from organisms listed in Table 1). The panel can also be used for training and evaluation of operator proficiency.

WARNINGS AND PRECAUTIONS:

- NATtrol™ inactivation was carried out on microorganism stocks used to formulate the panel members. The inactivation was verified in a standard microbiological growth protocol.
- This panel contains inactivated microorganisms and materials of human and animal origin. Safe practices suggest that the controls be considered potentially infectious and to use Universal Precautions when handling.
- Refer to CDC guidelines and local regulations for handling and disposal.
- The matrix used in the manufacture of this product is treated with 0.09% sodium azide. It was manufactured from Human Serum Albumin that have been tested and found to be non-reactive at the donor level for HIV-1/HIV-2 Antibody, HBsAg and HCV Antibody by FDA licensed donor screening test methods. All materials are also tested for HIV-1 and HCV by FDA approved Nucleic Acid Test (NAT) methods.
- Heat inactivated Fetal Bovine Serum used in the manufacture of this product meet applicable USDA requirements for abattoir sourced animals, traceability and country of origin. The materials were collected at USDA licensed establishments or legally imported from countries recognized by the USDA as negligible or controlled for risk for Bovine Spongiform Encephalopathy (BSE) and other exotic disease agents. Donor animals were inspected ante and post mortem at the abattoir as required by the USDA.
- Do not use past the expiration date on the label.
- To avoid cross-contamination, use separate pipette tips for all materials.

RECOMMENDED STORAGE:

- NATtrol™ Respiratory Verification Panel 2.1 should be stored at 2-8°C.

LIMITATION:

- **FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES**
- Use in accordance with local, state, federal, and accreditation requirements.
- This product is not intended to replace the manufacturer's controls provided with the assay.

EXPECTED RESULTS:

- This panel has been tested with the BioFire® Respiratory Panel 2.1 (RP2.1) assay and provides all expected results for the panel members listed in Table 1. This panel has also been tested on the BioFire® Respiratory Panel 2 (RP2) and provides all expected results.
- Each laboratory must evaluate the product and establish their own performance criteria.
- The table shown below is for informational purposes only.

TABLE 1: PANEL MEMBERS

Panel Member	Strain
Adenovirus 1	N/A
Adenovirus 3	N/A
Adenovirus 31	N/A
<i>B. parapertussis</i>	A747
<i>B. pertussis</i>	A639
<i>C. pneumoniae</i>	CWL-029
Coronavirus 229E	N/A
Coronavirus HKU-1	Recombinant ¹
Coronavirus NL63	N/A
Coronavirus OC43	N/A
Influenza A H1N1pdm	A/NY/02/09 ²
Influenza AH1	A/New Caledonia/20/99
Influenza AH3	A/Brisbane/10/07
Influenza B	B/Florida/02/06
<i>M. pneumoniae</i>	M129
Metapneumovirus 8	Peru6-2003 ³
Parainfluenza 1	N/A
Parainfluenza 2	N/A
Parainfluenza 3	N/A
Parainfluenza 4	N/A
Rhinovirus 1A	N/A
RSV A	N/A
SARS-CoV-2	USA-WA1/2020 ⁴
Negative Control	N/A








¹ This analyte only contains a short sequence of the viral genome therefore each laboratory must evaluate performance in their assay.

² Please note that although similar in nomenclature, **this is a 2009 H1N1 pandemic Influenza strain** and does NOT correlate with the seasonal 2009 Influenza strains found in the Fludb.org database. For reference, the NCBI Taxon IDs for the seasonal Influenza strains listed in the Fludb.org database are: A/New York/01/2009 (H1N1) - 666252; B/New York/01/2009 - 664512; A/New York/02/2009 (H1N1) - 666298; and A/New York/03/2009 (H3N2) - 659637.

³ This product is sold by ZeptoMetrix under license from Vironovative B. V. under patent applications, including U.S. Patent Applications 10/371,099 and 10/371,12 and any patents that issue from applications related to PCT/NL02/00040 and PCT/US03/05271.

⁴ This reagent was deposited by the Centers for Disease Control and Prevention and obtained through BEI Resources, NIAID, NIH: SARS-Related Coronavirus 2, Isolate USA-WA1/2020, NR-52281.

PINATRVP2.1-BIO
Revision: 04
Effective Date: 04/17/2024

	Catalog Number		Temperature Limitation
	Batch Code		Expiration Date
	For Research Use Only		Biological Risk
	Manufacturer		

PCA# 24-167
Page 1 of 1